



Human Complement Factor H (FH) Deficient Plasma (Sodium Citrate)

Catalog No. D700501

Lot No. FH0415RSC

Exp Date: 31MAR17

Reagent Details

- 1 vial of Deficient Human Plasma (1 mL, Lyophilized)
- The plasma was established as better than 98% deficient on average by ELISA testing.
- Albumin level serves as control for depletion process

Reconstitution

- Add 1 ml reagent grade water to produce a 1x deficient plasma

Warnings

- The Deficient Plasma is for **Research Use Only. Not for use in Diagnostic Procedures.**
- Variations between laboratories may be caused by technique differences. It is recommended that each laboratory establish its own acceptable ranges.
- Although all whole blood was deemed HIV and HBsAg negative, use universal precautions while testing.

Storage Condition

- Before reconstitution store at -20°C, or below.
- After reconstitution **aliquot to limit repeated freezing and thawing** and store at -60°C, or below, until the expiration date.

