



## **AssayMax Human Interleukin-8 (IL-8) ELISA Kit**

Catalog No. EI1008-1

### **Introduction**

CXCL8 or Interleukin-8 (IL-8) is a member of the CXC chemokine subfamily of cytokines. This basic heparin-binding protein precursor contains 99 amino acids and the mature functional protein comprises 72 amino acids (1). IL-8 is proinflammatory and primarily mediates the activation and migration of neutrophils from peripheral blood into the sites of inflammation, injury, or infection in the tissue (2). IL-8 interacts with two receptors, CXCR1 and CXCR2, to activate leukocytes. Upon activation, both receptors couple to G protein to mediate phosphoinositide-hydrolysis, intracellular  $\text{Ca}^{2+}$  mobilization, chemotaxis, and exocytosis. CXCR1 is specific for IL-8 and activates phospholipase D and mediates respiratory burst (3, 4). IL-8 is involved in a wide variety of physiological and pathological processes, including host defense against bacterial infection, bronchiolitis, arteriosclerosis, autoimmune disorders of skin, bones, and joints, and angiogenesis-dependent disorders such as rheumatoid arthritis, tumor growth, and wound repair (5, 6).

### **Principal of the Assay**

The AssayMax Human IL-8 ELISA kit is designed for detection of IL-8 in human plasma, serum, tissue extracts and cell culture supernatants. This assay employs a quantitative sandwich enzyme immunoassay technique that measures IL-8 in less than 5 hours. A polyclonal antibody specific for human IL-8 has been pre-coated onto a microplate. IL-8 in standards and samples is sandwiched by the immobilized antibody and a biotinylated polyclonal antibody specific for human IL-8, which is recognized by a streptavidin-peroxidase conjugate. All unbound material is then washed away and a peroxidase enzyme substrate is added. The color development is stopped and the intensity of the color is measured.

### **Caution and Warning**

- **Prepare all reagents (working diluent buffer, wash buffer, standards, biotinylated-antibody, and SP conjugate) as instructed, prior to running the assay.**
- **Prepare all samples prior to running the assay. The dilution factors for the samples are suggested in this protocol. However, the user should determine the optimal dilution factor.**
- **Spin down the SP conjugate vial and the biotinylated-antibody vial before opening and using contents.**
- This kit is for research use only.
- The kit should not be used beyond the expiration date.
- The Stop Solution is an acid solution

## Reagents

- **IL-8 Microplate:** A 96 well polystyrene microplate (12 strips of 8 wells) coated with a polyclonal antibody against IL-8.
- **Sealing Tapes:** Each kit contains 3 pre-cut, pressure-sensitive sealing tapes that can be cut to fit the format of the individual assay.
- **IL-8 Standard:** Human IL-8 in a buffered protein base (4 ng, lyophilized).
- **Biotinylated IL-8 Antibody (100x):** A 100-fold concentrated biotinylated polyclonal antibody against IL-8 (80 µl).
- **MIX Diluent Concentrate (10x):** A 10-fold concentrated buffered protein base (30 ml).
- **Wash Buffer Concentrate (20x):** A 20-fold concentrated buffered surfactant (30 ml, 2 bottles).
- **Streptavidin-Peroxidase Conjugate (SP Conjugate):** A 100-fold concentrate (80 µl).
- **Chromogen Substrate:** A ready-to-use stabilized peroxidase chromogen substrate tetramethylbenzidine (8 ml).
- **Stop Solution:** A 0.5 N hydrochloric acid to stop the chromogen substrate reaction (12 ml).

## Storage Condition

- Store components of the kit at 2-8°C or -20°C upon arrival up to the expiration date.
- Store SP Conjugate and Biotinylated Antibody at -20°C
- Store Microplate, Diluent Concentrate (10x), Wash Buffer, Stop Solution, and Chromogen Substrate at 2-8°C
- Opened unused microplate wells may be returned to the foil pouch with the desiccant packs. Reseal along zip-seal. May be stored for up to 1 month in a vacuum desiccator.
- Diluent (1x) may be stored for up to 1 month at 2-8°C.
- Store Standard at 2-8°C before reconstituting with Diluent and at -20°C after reconstituting with Diluent.

## Other Supplies Required

- Microplate reader capable of measuring absorbance at 450 nm.
- Pipettes (1-20 µl, 20-200 µl, 200-1000 µl and multiple channel).
- Deionized or distilled reagent grade water.

## Sample Collection, Preparation and Storage

- **Plasma:** Collect plasma using one-tenth volume of 0.1 M sodium citrate as an anticoagulant. Centrifuge samples at 2000 x g for 10 minutes and assay. Store samples at -20°C or below. Avoid repeated freeze-thaw cycles. (EDTA or Heparin can also be used as anticoagulant.)
- **Serum:** Samples should be collected into a serum separator tube. After clot formation, centrifuge samples at 2000 x g for 10 minutes. Remove serum and assay. Store serum at -20°C or below. Avoid repeated freeze-thaw cycles.
- **Cell Culture Supernatants:** Centrifuge cell culture media at 2000 x g for 10 minutes to remove debris. Collect supernatants and assay. Store samples at -20°C or below. Avoid repeated freeze-thaw cycles.
- **Tissue:** Extract tissue samples with 0.1M Tris-buffered saline (pH7.4) containing 0.5% Triton X-100 and centrifuge at 14000 x g for 30 min. Collect the supernatant and measure the protein concentration. Dilute the tissue extract 1:2 into MIX Diluent and assay. Freeze the remaining extract at -20°C or below. Avoid repeated freeze-thaw cycles.

## Reagent Preparation

- Freshly dilute all reagents and bring all reagents to room temperature before use.
- **MIX Diluent Concentrate (10x):** If crystals have formed in the concentrate, mix gently until the crystals have completely dissolved. Dilute the MIX Diluent Concentrate 1:10 with reagent grade water. Store for up to 1 month at 2-8°C.
- **Standard Curve:** Reconstitute the 4 ng of human IL-8 Standard with 4 ml of MIX Diluent to generate a standard solution of 1 ng/ml. Allow the standard to sit for 10 minutes with gentle agitation prior to making dilutions. Prepare duplicate or triplicate standard points by serially diluting the IL-8 standard solution twofold with equal volume of MIX Diluent to produce 0.5, 0.25, 0.125, 0.063, 0.031, and 0.016 ng/ml. MIX Diluent serves as the zero standard (0 ng/ml). Any remaining solution should be frozen at -20°C.

Standard Point	Dilution	[IL-8] (ng/ml)
P1	Standard (1 ng/ml)	1.000
P2	1 part P1 + 1 part MIX Diluent	0.500
P3	1 part P2 + 1 part MIX Diluent	0.250
P4	1 part P3 + 1 part MIX Diluent	0.125
P5	1 part P4 + 1 part MIX Diluent	0.063
P6	1 part P5 + 1 part MIX Diluent	0.031
P7	1 part P6 + 1 part MIX Diluent	0.016
P8	MIX Diluent	0.000

- **Biotinylated IL-8 Antibody (100x):** Spin down the antibody briefly and dilute the desired amount of the antibody 1:100 with MIX Diluent. Any remaining solution should be frozen at -20°C.
- **Wash Buffer Concentrate (20x):** If crystals have formed in the concentrate, mix gently until the crystals have completely dissolved. Dilute the Wash Buffer Concentrate 1:20 with reagent grade water.
- **SP Conjugate (100x):** Spin down the SP Conjugate briefly and dilute the desired amount of the conjugate 1:100 with MIX Diluent. Any remaining solution should be frozen at -20°C.

## Assay Procedure

- Prepare all reagents, working standards and samples as instructed. Bring all reagents to room temperature before use. The assay is performed at room temperature (20-30°C).
- Remove excess microplate strips from the plate frame and return them immediately to the foil pouch with desiccant inside. Reseal the pouch securely to minimize exposure to water vapor and store in a vacuum desiccator.
- Add 50 µl of Standard or sample per well. Cover wells with a sealing tape and incubate for two hours. Start the timer after the last sample addition.
- Wash five times with 200 µl of Wash Buffer manually. Invert the plate each time and decant the contents; hit it 4-5 times on absorbent paper towel to completely remove the liquid. If using a machine wash six times with 300 µl of Wash Buffer and then invert the plate, decant the contents; hit it 4-5 times on absorbent paper towel to completely remove the liquid.
- Add 50 µl of Biotinylated IL-8 Antibody to each well and incubate for two hours.
- Wash a microplate as described above.
- Add 50 µl of Streptavidin-Peroxidase Conjugate per well and incubate for 30 minutes. Turn on the microplate reader and set up the program in advance.
- Wash a microplate as described above.


- Add 50 µl of Chromogen Substrate per well and incubate for approximately 15 minutes or till the optimal blue color density develops. Gently tap the plate to ensure thorough mixing and break the bubbles in the well with pipette tip.
- Add 50 µl of Stop Solution to each well. The color will change from blue to yellow.
- Read the absorbance on a microplate reader at a wavelength of 450 nm immediately. If wavelength correction is available, subtract readings at 570 nm from those at 450 nm to correct optical imperfections. Otherwise, read the plate at 450 nm only. Please note that some unstable black particles may be generated at high concentration points after stopping the reaction for about 10 minutes, which will reduce the readings

## Data Analysis

- Calculate the mean value of the duplicate or triplicate readings for each standard and sample.
- To generate a Standard Curve, plot the graph using the standard concentrations on the x-axis and the corresponding mean 450 nm absorbance on the y-axis. The best-fit line can be determined by regression analysis using log-log or four-parameter logistic curve-fit.
- Determine the unknown sample concentration from the Standard Curve and multiply the value by the dilution factor.

## Standard Curve

- The curve is provided for illustration only. A standard curve should be generated each time the assay is performed.



## Performance Characteristics

- The minimum detectable dose of IL-8 is typically ~ 0.01 ng/ml.
- Intra-assay and inter-assay coefficients of variation were 4.9% and 7.3% respectively.
- This assay recognizes both natural and recombinant human IL-8.

## Linearity

Sample Dilution	Average Percentage of Expected Value	
	Plasma	Serum
No dilution	96%	99%
1:2	97%	97%
1:4	99%	95%

## Recovery

Standard Added Value	0.03 – 0.3 ng/ml
Recovery %	85-117%
Average Recovery %	98 %

## Cross-Reactivity

Species	% Cross Reactivity
Bovine	5%
Mouse	5%
Rabbit	5%
Rat	1%
Swine	5%
Canine	5%
Monkey	60%

## References

- (1) Matsushima K et al. (1988) J. Exp. Med. 267(6): 1883-1893
- (2) Harada A et al. (1996) Mol. Med. Today 2: 482-489
- (3) L'Heureux GP et al. (1995) Blood 85: 522-531
- (4) Jones SA et al. (1996) Proc. Natl. Acad. Sci. USA 93: 6682-6686
- (5) Koch AE et al. (1995) Science 268(5209): 447 –448
- (6) Jeremy H et al. (2001) Am J Hum Genet. 69(2): 413–419

Version 1.9