

Mouse Interleukin 12, II12 ELISA Kit

Cat. No.: DEIA118

Lot. No.: (See product label)

Pkg.Size: 10 Plates

INTENDED USE

Mouse IL-12 ELISA development kit contains the key components required for the quantitative measurement of natural and/or recombinant IL-12 in a sandwich ELISA format within the range of 32–2,000 pg/ml. Using the ELISA protocol described below, this kit provides sufficient reagents to assay IL-12 in approximately 1,000 ELISA plate wells.

DESCRIPTION

Interleukin-12 is a disulfide-linked heterodimeric protein (p70), composed of two subunits, p35 and p40, which are encoded by two different genes. Accumulating data indicate that p40 secretion precedes that of IL-12 expression. In addition, to its ability to covalently bind to p35 to form IL-12, p40 can bind to p19 to form IL-23.

RECONSTITUTION & STORAGE


<u>Capture Antibody:</u> 100 µg of antigen-affinity purified goat anti-IL-12. Centrifuge vial prior to opening. Reconstitute in 1.0 ml sterile water for a concentration of 100 µg/ml. Following reconstitution the Capture antibodies may be stored at 2-8 °C for up to 6 months. For long term storage it is recommended to aliquot into working volumes and store at -70 °C in a manual defrost freezer.

<u>Detection Antibody:</u> 25 μg of biotinylated antigen-affinity purified goat anti-IL-12. Centrifuge vial prior to opening. Reconstitute in 0.25 ml sterile water for a concentration of 100 μg/ml. Following reconstitution the Detection antibodies may be stored at 2-8°C for up to 6 months. For long term storage it is recommended to aliquot into working volumes and store at -70°C in a manual defrost freezer.

Mouse IL-12 Standard: 1 μg of recombinant IL-12. Centrifuge vial prior to opening. Reconstitute in 1.0 ml sterile water for a concentration of 1 μg/ml. The Standard may be stored at 2-8 $^{\circ}$ C for one month or aliquoted and stored at -70 $^{\circ}$ C for up to three months in a manual defrost freezer.

<u>UltraAvidin-HRP Conjugate:</u> 40 μ l vial. Upon receipt, UltraAvidin-HRP conjugate should be stored at 2-8 $^{\circ}$ C, **DO NOT FREEZE**

<u>TMB Liquid Substrate</u>: Aspirate and wash plate 4 times. Add 100 μ l of TMB HRPO Microwell Substrate Standard Kinetic One Component "Ready Use" to each well. Incubate at room temperature and monitorcolor development with an ELISA plate reader at 450 nm with wavelength correction set at 540 nm or 570 nm. The reaction may be stopped after 15 – 20 minutes by adding 100 μ l of 2 M sulfuric acid to each well.



Crystal structure of human IL-12

ANALYTE GENE INFORMATION

Gene Name: Il12a interleukin 12a [Mus musculus]

Official Symbol: II12a

Synonyms: Il12a; interleukin 12a; p35; Ll12a; Il-12a; IL-12p35; MGC151228; MGC151232; interleukin 12a p35

subunit; interleukin 12a *GeneID*: 16159

mRNA Refseq: NM 001159424
Protein Refseq: NP 001152896

UniProt ID: P43431

Chromosome Location: 3 E1; 3 37.0 cM

Function: cytokine activity; growth factor activity; interleukin-12 beta subunit binding; interleukin-12 receptor bind-

ing

RECOMMENDED MATERIALS

- ELISA microplates;
- BSA:
- Stop Solution 2 M Sulfuric Acid;
- Dulbecco's PBS [10x]

RECOMMENDED SOLUTIONS

<u>PBS:</u> Dilute 10xPBS to 1xPBS, pH 7.2 in sterile water. <u>Wash Buffer:</u> 0.05% Tween-20 in PBS.

<u>Block Buffer:</u> 1.0% BSA in PBS. **Diluent:** 1.0% BSA in PBS.

PLATE PREPARATION

- 1. Dilute to 2.0 μ g/ml of capture antibody and immediately add 100 μ l to each ELISA plate well. Seal the plate and incubate overnight at room temperature.
- 2. Aspirate the wells to remove liquid and wash the plate 4 times using 300µl of wash buffer per well.
- 3. After the last wash invert plate to remove residual buffer and blot on paper towel.
- 4. Add 300µl block buffer to each well. Incubate for at least 1 hour at room temperature.
- 5. Aspirate and wash plate 4 times.



ELISA PROTOCOL

<u>Standard/Sample:</u> Dilute standard from 2,000 pg/ml to zero in diluent. Immediately add 100 µl of standard or sample to each well in duplicate and incubate at room temperature for at least 2 hours.

<u>**Detection:**</u> Aspirate and wash plate 4 times. Dilute a portion of detection antibody in diluent to a concentration of 50 ng/ml. Add 100 μ l per well and incubate at room temperature for 2 hours.

<u>UltraAvidin-HRP Conjugate:</u> Aspirate and wash plate 4 times. Dilute 3.0 μ l of avidin-HRP conjugate 1:5,000 (this dilution factor may require some optimization) in diluent for a total volume of 15 ml. Add 100 μ l per well and incubate at room temperature for 30 minutes.

TMB Liquid Substrate: Add 100 μl of TMB HRPO Microwell Substrate Standard Kinetic One Component "Ready Use" to each well. Incubate at room temperature for 20-30 minutes and monitor color development with an ELISA plate reader at 450 nm with wavelength correction set at 540 nm or 570 nm. Avoid placing plates in direct light.

<u>Stop Solution:</u> The reaction may be stopped after 15-20 minutes by adding $100~\mu l$ of 2~M sulfuric acid to each well. All eye, hand, face, and clothing protection should be used when working with sulfuric acid.


REFENRENCES

- 1. Airoldi I, Guglielmino R, Carra G, et al. (2002). "The interleukin-12 and interleukin-12 receptor system in normal and trans-formed human B lymphocytes.". Haematologica 87 (4): 434–42.
- 2. Kim JH, Loveland JE, Sitz KV, et al. (1997). "Expansion of restricted cellular immune responses to HIV-1 envelope by vaccination: IL-7 and IL-12 differentially augment cellular proliferative responses to HIV-1.". Clin. Exp. Immunol. 108 (2): 243–50.

CROSS REACTIVITY

When tested at 50 ng/ml the following antigen exhibited less than 5% cross reactivity: Human IL-12, IFNγ, IL-10 and IL-12p40. Mouse IFNγ and IL-10.

TYPICAL STANDARD CURVE



The standard curve is provided as an example only. A standard curve should be prepared with each assay.