

## Human M2 type mitochondria Antibody ELISA Kit

Cat. No.:DEIA1224

Pkg.Size:96T

### Intended use

The Human M2 type mitochondria Antibody ELISA kit is designed to detect and quantify the level of human M2 type mitochondria Antibody in cell culture supernatant, serum, plasma and tissue.

### General Description

In cell biology, a mitochondrion (plural mitochondria) is a membrane-enclosed organelle found in most eukaryotic cells. These organelles range from 0.5 to 10 micrometers ( $\mu\text{m}$ ) in diameter. Mitochondria are sometimes described as cellular power plants because they generate most of the cell's supply of adenosine triphosphate (ATP), used as a source of chemical energy. In addition to supplying cellular energy, mitochondria are involved in a range of other processes, such as signaling, cellular differentiation, cell death, as well as the control of the cell cycle and cell growth. Mitochondria have been implicated in several human diseases, including mitochondrial disorders and cardiac dysfunction and may play a role in the aging process.

### Principle Of The Test

The quantitative immunoenzymatic determination of antibodies against M2 type mitochondria antigen is based on the ELISA (Enzyme-linked Immunosorbent Assay) technique. Microtiter strip wells are precoated with M2 type mitochondria antigen to bind corresponding antibodies of the specimen. After washing the wells to remove all unbound sample material, and goat anti-human IgG antibodies (anti-IgG) conjugated to horseradish peroxidase (HRP) are added. This conjugate binds to antigen-antibody complexes. The immune complex formed by the bound conjugate is visualized by adding Tetramethylbenzidine (TMB) substrate which gives a blue reaction product. The intensity of this product is proportional to the amount of M2 type mitochondria antigen - specific antibodies in the specimen. Sulphuric acid is added to stop the reaction. This produces a yellow endpoint colour. Absorbance at 450 nm is read using an ELISA microwell plate reader.

### Reagents And Materials Provided

Human M2 type mitochondria Ag Coated Wells, 96-well polystyrene microplate (12 strips of 8 wells): 1 plate

Standard 1(800 RU/ml), 1 mL per bottle: 1 bottle

Standard 2(200 RU/ml), 1 mL per bottle: 1 bottle

Standard 3(100 RU/ml), 1 mL per bottle: 1 bottle

Standard 4(50 RU/ml), 1 mL per bottle: 1 bottle

Standard 5(25 RU/ml), 1 mL per bottle: 1 bottle

Positive control, 1 mL per bottle: 1 bottle

Negative control, 1 mL per bottle: 1 bottle

HRP-Conjugate Reagent, 15 ml per bottle with preservatives: 1 bottle

Assay Solution, 100 mL per bottle with preservatives: 1 bottle

Wash Solution Concentrate (10 $\times$ ), 100 mL with preservatives: 1 bottle

Chromogen Solution A, 10 mL per bottle: 1 bottle

Chromogen Solution B, 10 mL per bottle: 1 bottle

Stop Solution, 10 mL per bottle: 1 bottle

### Materials Required But Not Supplied

**Creative Diagnostics. All rights reserved**

45-16 Ramsey Road Shirley, NY 11967, USA

Tel: 631-624-4882 · Fax: 631-614-7828

E-mail: [info@creative-diagnostics.com](mailto:info@creative-diagnostics.com)

[www.creative-diagnostics.com](http://www.creative-diagnostics.com)

1. A standard ELISA plate reader for absorbance at 450 nm.
2. Calibrated adjustable precision pipettes (single channel and multi channel), preferably with disposable plastic tips.
3. Distilled or deionized water.
4. Plate washer: automated or manual (squirt bottle, manifold dispenser, etc.).
5. Data analysis and graphing software or graph paper.
6. Polypropylene tubes.
7. Graduated cylinders and calibrated beakers in various sizes.

## Storage

Store all reagents at 2 to 8°C.

## Specimen Collection And Handling

1. Sample Collection: Blood collected by venipuncture should be allowed to clot naturally and completely – the serum/plasma must be separated from the clot as early as possible as to avoid hemolysis of the RBC and the volume should >50 µl.
2. Storage: Store samples at 2-8 °C. Samples not required for assaying within 3 days should be stored frozen (-20°C or lower). Multiple freeze-thaw cycles should be avoided.

## Reagent Preparation

1. Allow the reagents and samples to reach room temperature (18-30°C) for at least 15-30 minutes and mix gently.
2. Wash Solution: Make a 1:10 dilution of Wash Solution Concentrate (10X) with deionized or distilled water in a clean plastic tube as needed. Label as Wash Solution. Store both the concentrate and the Wash Solution in the refrigerator. The diluted buffer should be used within one week.
3. Diluting Sample: Dilute each sample 1:200 with sample diluent. The diluted buffer should be used within 6 hours.

## Assay Steps

1. Determine the number of strips needed for the assay and remove excess microplate strips from the plate frame, return them to the foil pouch, and reseal.
2. Adding Sample:  
Quantitative test: Add 100µl of samples and 100µl Standard, Positive and Negative controls into their respective wells.  
Qualitative test: Add 100µl of samples and 100µl Standard 5, Positive and Negative controls into their respective wells. Seal the plate with Plate Covers and incubate at 37°C for 30min. Note: Use a separate disposal pipette tip for each specimen, Negative and Positive Controls as to avoid cross-contamination.
3. Aspirate each well and wash 4 times. Wash by filling each well with Wash Solution (300 µL) using a multi-channel pipette, squirt bottle, manifold dispenser or autowasher. Allow the Wash Solution to soak for about 10-20 seconds before aspiration. After the last wash step, the plate is inverted and tapped dry on absorbent pad or paper towel. A thorough washing of the plate is extremely important to reduce background. (Prepare Wash Solution according to REAGENT PREPARATION.)
4. Add 100 µL of Streptavidin-HRP Solution to each well. Cover with a new adhesive strip. Incubate for 30 minutes at 37°C.
5. Aspirate each well and wash according to Step 3.
6. Dispense 50µl of Chromogen A and 50µl Chromogen B solution into each well including the Blank and mix by tapping the plate gently. Incubate the plate at 37°C for 15 minutes avoiding light. The enzymatic reaction between the Chromogen A/B solutions produces blue color in Positive control and SSA antibody Positive sample wells.
7. Stopping Reaction: Using a multichannel pipette or manually, add 50µl Stop solution into each well and mix gently. Intensive yellow color develops in Positive control and SSA antibody positive sample wells.
8. Measuring the Absorbance: Calibrate the plate reader with the Blank well and read the absorbance at 450nm. If a dual filter instrument is used, set the reference wavelength at 630nm. Calculate the Cut-off value and evaluate the results. (Note: read the

absorbance within 5 minutes after stopping the reaction)

## Quality Control

The test results are valid if the Quality Control criteria are verified. It is recommended that each laboratory must establish appropriate quality control system with quality control material similar to or identical with the sample being analyzed.

1. The OD value of the Blank well, which contains only Chromogens and Stop solution, is less than 0.080 at 450 nm.
2. The OD value of the Positive control must be equal to or greater than 0.800 at 450/630nm or at 450nm after blanking.
3. The OD value of the Negative control must be less than 0.100 at 450/630nm or at 450nm after blanking.

## Calculation

Average the duplicate readings for each standard and sample and subtract the average zero standard optical density.

**Manual Plotting:** Plot on graph paper the absorbance of the standards against the standard concentration. Known concentrations of Human M2 type mitochondria are plotted on the X-axis and the corresponding absorbances on the Y-axis. The standard curve should result in a straight line that shows a direct relationship between Human M2 type mitochondria concentrations and the corresponding absorbances. The concentration of Human M2 type mitochondria in samples may be determined by plotting the sample absorbances on the Y-axis, then drawing a horizontal line to intersect with the standard curve. At the point of intersection, extend a vertical line to the abscissa and read the corresponding Human M2 type mitochondria concentration. Note: Samples producing signals greater than that of the highest standard should be diluted in Assay Solution and reanalyzed. Multiply the measured concentration by the appropriate dilution factor.

**Plate Reader:** An alternative approach is to use an ELISA curve fitting software. A linear curve plot is expected to produce the best fit of the resulting sample concentrations.

## Interpretation of Results

### 1. Quality

sample/control=O.D. with sample/O.D. with Standard 5

Negative Results (sample/control<1) : samples giving absorbance less than the O.D. with Standard 5 are negative for this assay, which indicates that no antibodies to M2 type mitochondria antigen have been detected with this anti-M2 type mitochondria antibody ELISA kit. Therefore, there are no serological evidences for recent infections with M2 type mitochondria and the patient is probably not infected with the virus.

Positive Results (sample/control≥1) : samples giving an absorbance greater than, or equal to the O.D. with Standard 5 are initially reactive, which indicates that antibodies to M2 type mitochondria antigen have probably been detected with this anti-M2 type mitochondria antibody ELISA kit. Any reactive samples should be retested in duplicates. Repeatedly reactive samples could be considered positive for anti-M2 type mitochondria antibody. Positive results with anti-M2 type mitochondria antibody detection indicate possible recent infection with M2 type mitochondria.

Borderline (sample/control =0.9-1.1): Samples with absorbance to O.D. with Standard 5 between 0.9 and 1.1 are considered borderline samples and retesting of these samples in duplicates is recommended. Repeatedly positive samples can be considered positive for anti-M2 type mitochondria antibody.

The result from this M2 type mitochondria should not be used alone to establish the infection state.

### 2. Quantify

Sample with concentration < 25 RU/ml are negative for this assay.

Sample with concentration ≥25 RU/ml are positive for this assay.

Sample with concentration =25 RU/ml-27 RU/ml are considered borderline samples and retesting of these samples in duplicates is recommended. Repeatedly positive samples can be considered positive for anti-M2 type mitochondria antibody.

## Sensitivity

The sensitivity is defined as the probability of the assay of scoring positive in the presence of the specific analyte. It is  $\geq 90\%$ .

## Specificity

The specificity is defined as the probability of the assay of scoring negative in the absence of the specific analyte. It is  $\geq 90\%$ .

## Reproducibility

Intra-Assay:  $CV \leq 15\%$

Inter-Assay:  $CV \leq 15\%$

## Precautions

1. The Stop solution (2M H<sub>2</sub>SO<sub>4</sub>) is a strong acid. Corrosive. Use it with appropriate care. Wipe up spills immediately or wash with water if come into contact with the skin or eyes. Proclin 300 used as a preservative can cause sensation of the skin.
2. All specimens from human origin should be considered as potentially infectious. However, there is no analytical method that can assure that infectious agents in the specimens or reagents are completely absent. Therefore, handle reagents and specimens with extreme caution as if capable of transmitting infectious diseases. Strict adherence to GLP (Good Laboratory Practice) regulations can ensure the personal safety. Never eat, drink, smoke, or apply cosmetics in the assay laboratory.
3. Make sure that all reagents are within the validity indicated on the kit box and are of the same lot. Never use reagents beyond the expiry date stated on reagents labels or on the kit box.
4. Do not exchange reagents from different lots, or use reagents from other commercially available kits. The components of the kit are precisely matched as to achieve optimal performance during testing.

## Limitations

This is a qualitative assay and the results cannot be use to measure antibodies concentrations.