Human Interleukin 17F, IL17F ELISA Kit

Cat. No.: DEAI127
Lot. No.: (See product label)
Pkg.Size: 10 Plates

INTENDED USE

Human IL-17F ELISA development kit contains the key components required for the quantitative measurement of natural and/or recombinant IL-17F in a sandwich ELISA format within the range of 32–2,000 pg/ml. Using the ELISA protocol described below, this kit provides sufficient reagents to assay IL -17F in approximately 1,000 ELISA plate wells.

DESCRIPTION

IL-17F is a homodimeric protein that is a member of the IL-17 family of cytokines produced by activated T-cells and monocytes. IL-17F is expressed by activated T cells and can stimulate production of other cytokines such as IL-6, IL-8 and granulocyte colony-stimulating factor, and can regulate cartilage matrix turnover. IL-17F is an important regulator of inflammatory responses that function differently than IL-17 in immune responses and diseases.

RECONSTITUTION & STORAGE

Capture Antibody: 100 µg of antigen-affinity purified goat anti-IL-17F. Centrifuge vial prior to opening. Reconstitute in 1.0 ml sterile water for a concentration of 100 µg/ml. Following reconstitution the Capture antibodies may be stored at 2-8℃ for up to 6 months. For long term storage it is recommended to aliquot into working volumes and store at -70℃ in a manual defrost freezer.
Detection Antibody: 100 µg of biotinylated antigen-affinity purified goat anti-IL-17F. Centrifuge vial prior to opening. Reconstitute in 1.0 ml sterile water for a concentration of 100 µg/ml. Following reconstitution the Detection antibodies may be stored at 2-8℃ for up to 6 months. For long term storage it is recommended to aliquot into working volumes and store at -70℃ in a manual defrost freezer.
Human IL-17F Standard: 1 µg of recombinant IL-17F. Centrifuge vial prior to opening. Reconstitute in 1.0 ml sterile water for a concentration of 1 µg/ml. The Standard may be stored at 2 -8℃ for one month or aliquoted and stored at -70℃ for up to three months in a manual defrost freezer.
UltraAvidin-HRP Conjugate: 40 µl vial. Upon receipt, UltraAvidin-HRP conjugate should be stored at 2-8℃, **DO NOT FREEZE.**
TMB Liquid Substrate: Aspirate and wash plate 4 times. Add 100 µl of TMB HRPO Microwell Substrate Standard Kinetic One Component "Ready Use" to each well. Incubate at room temperature and monitorcolor development with an ELISA plate reader at 450 nm with wavelength correction set at 540 nm or 570 nm. The reaction may be stopped after 15 – 20 minutes by adding 100 µl of 2 M sulfuric acid to each well.

ANALYTE GENE INFORMATION

Gene Name: IL17F interleukin 17F [*Homo sapiens*]
Official Symbol: IL17F
Synonyms: IL17F; interleukin 17F; ML1; ML-1; IL-17F; cytokine ML-1; Interleukin-24; leukin-17F
GeneID: 112744
mRNA Refseq: NM_052872
Protein Refseq: NP_443104
MIM: 606496
UniProt ID: Q96PD4
Chromosome Location: 6p12
Function: cytokine activity; cytokine binding; protein homodimerization activity

RECOMMENDED MATERIALS

- ELISA microplates;
- BSA;
- Stop Solution 2 M Sulfuric Acid;
- Dulbecco's PBS [10x]

RECOMMENDED SOLUTIONS

PBS: Dilute 10xPBS to 1xPBS, pH 7.2 in sterile water.
Wash Buffer: 0.05% Tween-20 in PBS.
Block Buffer: 1.0% BSA in PBS.
Diluent: 1.0% BSA in PBS.

PLATE PREPARATION

1. Dilute to 2.0 µg/ml of capture antibody and immediately add 100 µl to each ELISA plate well. Seal the plate and incubate overnight at room temperature.
2. Aspirate the wells to remove liquid and wash the plate 4 times using 300µl of wash buffer per well.
3. After the last wash invert plate to remove residual buffer and blot on paper towel.
4. Add 300µl block buffer to each well. Incubate for at least 1 hour at room temperature.
5. Aspirate and wash plate 4 times.

45-16 Ramsey Road Shirley, NY 11967, USA
Tel: 631-624-4882 ·Fax:631-614-7828
E-mail: info@creative-diagnostics.com
www.creative-diagnostics.com

ELISA PROTOCOL

Standard/Sample: Dilute standard from 2,000 pg/ml to zero in diluent. Immediately add 100 µl of standard or sample to each well in duplicate and incubate at room temperature for at least 2 hours.
Detection: Aspirate and wash plate 4 times. Dilute a portion of detection antibody in diluent to a concentration of 50 ng/ml. Add 100 µl per well and incubate at room temperature for 2 hours.
UltraAvidin-HRP Conjugate: Aspirate and wash plate 4 times. Dilute 3.0 µl of avidin-HRP conjugate 1:5,000 (this dilution factor may require some optimization) in diluent for a total volume of 15 ml. Add 100µl per well and incubate at room temperature for 30 minutes.
TMB Liquid Substrate: Add 100 µl of TMB HRPO Microwell Substrate Standard Kinetic One Component "Ready Use" to each well. Incubate at room temperature for 20-30 minutes and monitor color development with an ELISA plate reader at 450 nm with wavelength correction set at 540 nm or 570 nm. Avoid placing plates in direct light.
Stop Solution: The reaction may be stopped after 15 – 20 minutes by adding 100 µl of 2 M sulfuric acid to each well. All eye, hand, face, and clothing protection should be used when working with sulfuric acid.

REFENRENCES

1. Neote K *et al.* (1998) *J Exp Med.* 187: 2009.
2. Losey J *et al.* (2007) *Eur Neurol.* 58: 228.
3. Barr ML *et al.* (2007) *Mol Cell Biochem.* 296:1.
4. Jacek Losy *et al.* (2007) *Eur Neurol.* 58: 228.

CROSS REACTIVITY

When tested at 50 ng/ml the following antigen exhibited less than 5% cross reactivity: Human IL-17A, G-CSF, IFNγ, IL-1α, IL-4, IL-5, IL-6, IL-8, IL-10, IL-12, IL-13, IL-16, IL-17B, IL-17D, IL-17E and TNF-α. Murine IL-17 and IL-17F.

TYPICAL STANDARD CURVE



The standard curve is provided as an example only. A standard curve should be prepared with each assay.

45-16 Ramsey Road Shirley, NY 11967, USA
Tel: 631-624-4882 ·Fax:631-614-7828
E-mail: info@creative-diagnostics.com
www.creative-diagnostics.com